

Name: Inner Mongolia Institute of Metrology and Testing

Address: No.147, Hulun South Road, Saihan District, Hohhot, Inner Mongolia, China

Registration No. CNAS L1645

Accreditation Criteria: ISO/IEC 17025:2017 and relevant requirements of CNAS

Effective Date: 2026-03-23 Expiry Date: 2030-11-06

CHINA NATIONAL ACCREDITATION SERVICE FOR CONFORMITY ASSESSMENT
SCHEDULE OF ACCREDITATION CERTIFICATE

SCHEDULE 5 ACCREDITED CALIBRATION AND MEASUREMENT CAPABILITY SCOPE

Note: The instruments with * represents onsite calibration can be performed.

No	Instrument	Measurand	Calibration Method	Range	Expanded Uncertainty (k=2)	Note	Effective Date
1、Geometric quantity							
1	Invar Bar-Coded Levelling Staffs	Length	Verification Regulation of Invar Bar-Coded Levelling Staffs JJG (mapping) 2102	(0~3) m	$U=5.3 \mu m$		



No. CNAS L1645